

## MT200 FCS Measurement

fluorescence correlation spectroscopy (FCS) is a quantitative measurement of the resolution and sensitivity of confocal microscopes. In this tutorial we follow molecular brightness, the count rate per molecule per excitation power, as a well suited parameter for comparison of instrument performance and quality of adjustment. The higher the better: if the system is not optimally aligned, the excitation and/or the detection will be less efficient and the molecular brightness will drop.

Copyright of this document belongs to PicoQuant GmbH. No parts of it may be reproduced, translated or transferred to third parties without written permission of PicoQuant GmbH. All information given here is reliable to our best knowledge. However, no responsibility is assumed for possible inaccuracies or omissions. Specifications and external appearances are subject to change without notice.



PicoQuant GmbH  
Rudower Chaussee 29 (IGZ)  
12489 Berlin  
Germany

P +49-(0)30-1208820-89  
F +49-(0)30-1208820-90  
info@picoquant.com  
www.picoquant.com